
Midmark M11 - v1.0.0
Total Cycles: 2792

Name: RITTER 1
Sterilizer ID: V2178198
01/18/2023 11:46 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 196.3 F 0.0
2:00 204.4 F 3.2
4:00 209.0 F 8.1
6:00 219.5 F 11.2
8:00 227.7 F 8.3
10:00 238.9 F 11.9
12:00 241.3 F 11.5
14:00 239.5 F 10.2
16:00 253.1 F 17.1
18:00 265.0 F 24.1

STERILIZING
mm:ss Degrees PSI
0:00 271.7 F 28.6
1:00 272.8 F 29.2
2:00 271.7 F 28.5
3:00 272.9 F 29.4
4:00 272.1 F 28.8

Min 271.7 F 28.5
Max 272.9 F 29.4

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 211.7 F 0.0
5:00 225.7 F 0.0
10:00 234.8 F 0.0
15:00 227.6 F 0.0
20:00 223.9 F 0.0
25:00 223.1 F 0.0
30:00 221.5 F 0.0

Drying Complete

FILLING: 1:35
HEATING: 19:51
STERILIZING: 4:00
VENTING: 3:13
DRYING: 30:00
TOTAL CYCLE: 00:58:39
CYCLE COMPLETE

01/18/2023 12:44 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____