
Midmark M11 - v1.0.0
Total Cycles: 3628

Name: RITTER 1
Sterilizer ID: V2178198
08/09/2023 02:26 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 203.2 F 0.0
2:00 192.7 F 4.3
4:00 196.2 F 8.9
6:00 220.4 F 7.5
8:00 227.6 F 9.2
10:00 239.0 F 10.4
12:00 241.3 F 11.9
14:00 240.0 F 10.6
16:00 253.4 F 17.3
18:00 265.0 F 24.2

STERILIZING
mm:ss Degrees PSI
0:00 271.2 F 28.6
1:00 272.2 F 29.2
2:00 271.3 F 28.6
3:00 272.2 F 29.3
4:00 271.5 F 28.6

Min 271.2 F 28.5
Max 272.5 F 29.4

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 212.1 F 0.0
5:00 216.9 F 0.0
10:00 235.9 F 0.0
15:00 240.6 F 0.0
20:00 232.7 F 0.0
25:00 223.7 F 0.0
30:00 222.8 F 0.0
Drying Complete

FILLING: 1:45
HEATING: 19:49
STERILIZING: 4:00
VENTING: 3:23
DRYING: 30:00
TOTAL CYCLE: 00:58:57
CYCLE COMPLETE

08/09/2023 03:24 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____