

-----  
North York, ON  
Canada  
-----

STATIM 2000 S203R750  
SN 500919I00007  
Unit # : 000  
WATER QUALITY  
0.2uS / 0.1ppm  
CYCLE NUMBER 001027  
9:46 07/04/2022  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 3:55  
PRESSURIZING 3:58  
STERILIZING 5:36  
136.0 C 221kPa 5:36  
136.3 C 224kPa 6:06  
136.4 C 224kPa 6:36  
136.2 C 223kPa 7:06  
136.3 C 224kPa 7:36  
136.4 C 225kPa 8:06  
136.4 C 225kPa 8:36  
136.4 C 224kPa 9:06  
136.4 C 224kPa 9:06  
VENTING 9:06  
AIR DRYING 10:06  
CYCLE COMPLETE 25:06  
-----

Digital Signature #  
6907D7B0E07E01F1  
-----