
Midmark M11 - v1.0.5
Total Cycles: 1665

Name: RITTER2
Sterilizer ID: V2313228
10/26/2022 05:03 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 175.4 F 0.0
2:00 177.6 F 3.8
4:00 192.3 F 9.0
6:00 216.0 F 6.7
8:00 231.5 F 10.8
10:00 233.0 F 8.6
12:00 238.6 F 10.1
14:00 247.7 F 14.8
16:00 260.5 F 22.0
18:00 272.0 F 29.7

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 29.3
1:00 272.5 F 29.9
2:00 271.3 F 29.4
3:00 272.4 F 30.0
4:00 271.6 F 29.3

Min 271.3 F 29.2
Max 272.6 F 30.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 204.4 F 0.0
5:00 213.3 F 0.0
10:00 228.7 F 0.0
15:00 229.7 F 0.0
20:00 229.5 F 0.0
25:00 229.2 F 0.0
30:00 227.5 F 0.0
Drying Complete

FILLING: 1:54
HEATING: 18:32
STERILIZING: 4:00
VENTING: 3:10
DRYING: 30:00
TOTAL CYCLE: 00:57:36
CYCLE COMPLETE

10/26/2022 06:00 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____