

-----  
North York, ON  
Canada  
-----

STATIM 2000 S203R750  
SN 500919I00007  
Unit # : 000  
WATER QUALITY  
0.0uS / 0.0ppm  
CYCLE NUMBER 001493  
10:52 04/10/2023  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 3:12  
PRESSURIZING 3:14  
STERILIZING 4:21  
136.1 C 221kPa 4:21  
136.2 C 222kPa 4:51  
136.4 C 225kPa 5:21  
136.4 C 221kPa 5:51  
136.4 C 225kPa 6:21  
136.4 C 224kPa 6:51  
136.1 C 222kPa 7:21  
136.4 C 225kPa 7:51  
136.4 C 225kPa 7:51  
VENTING 7:51  
AIR DRYING 8:39  
CYCLE COMPLETE 17:08

-----  
Digital Signature #  
215400BC82FCAD66  
-----