
Midmark M11 - v1.0.5
Total Cycles: 2994

Name: RITTER2
Sterilizer ID: V2313228
10/11/2023 09:36 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 130.1 F 0.0
2:00 137.1 F 0.7
4:00 143.0 F 3.2
6:00 168.6 F 7.0
8:00 194.3 F 11.7
10:00 212.7 F 5.6
12:00 229.0 F 10.7
14:00 228.1 F 7.0
16:00 240.9 F 13.0
18:00 240.6 F 11.5
20:00 253.8 F 18.1
22:00 265.1 F 25.0

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 29.4
1:00 272.6 F 29.7
2:00 271.8 F 29.9
3:00 272.1 F 29.5
4:00 272.3 F 30.1

Min 271.3 F 29.2
Max 272.7 F 30.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 204.1 F 0.0
5:00 230.0 F 0.0
10:00 200.0 F 0.0
15:00 200.2 F 0.0
20:00 201.7 F 0.0
25:00 200.5 F 0.0
30:00 202.8 F 0.0
Drying Complete

FILLING: 0:56
HEATING: 23:46
STERILIZING: 4:00
VENTING: 3:09
DRYING: 30:00
TOTAL CYCLE: 01:01:52
CYCLE COMPLETE

10/11/2023 10:37 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____