

-----  
Midmark M11 - v1.0.0  
Total Cycles: 1462

Name: RITTER 1  
Sterilizer ID: V2178198  
12/15/2021 04:15 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 170.6 F 0.0  
2:00 169.3 F 2.1  
4:00 173.3 F 5.4  
6:00 193.1 F 9.7  
8:00 215.2 F 7.3  
10:00 224.7 F 8.7  
12:00 235.4 F 12.0  
14:00 237.1 F 11.1  
16:00 235.3 F 9.1  
18:00 249.3 F 15.6  
20:00 260.6 F 22.0  
22:00 271.1 F 28.7

STERILIZING  
mm:ss Degrees PSI  
0:00 271.5 F 29.1  
1:00 272.4 F 29.8  
2:00 271.9 F 29.2  
3:00 271.5 F 29.3  
4:00 272.2 F 29.8

Min 271.3 F 28.9  
Max 272.4 F 29.8

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 206.0 F 0.0  
5:00 204.7 F 0.0  
10:00 218.1 F 0.0  
15:00 224.3 F 0.0  
20:00 227.3 F 0.0  
25:00 229.7 F 0.0  
30:00 229.1 F 0.0  
Drying Complete

FILLING: 1:40  
HEATING: 22:46  
STERILIZING: 4:00  
VENTING: 3:22  
DRYING: 30:00  
TOTAL CYCLE: 01:01:48  
CYCLE COMPLETE

12/15/2021 05:16 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_