
Midmark M11 - v1.0.5
Total Cycles: 4056

Name: RITTER2
Sterilizer ID: V2313228
10/25/2024 12:21 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 114.1 F 0.0
2:00 114.5 F 0.0
4:00 132.8 F 0.1
6:00 169.8 F 2.2
8:00 195.3 F 5.9
10:00 204.0 F 10.1
12:00 222.2 F 8.4
14:00 230.9 F 8.3
16:00 240.0 F 12.9
18:00 238.2 F 10.3
20:00 236.4 F 8.8
22:00 249.5 F 15.2
24:00 260.6 F 21.5
26:00 271.1 F 28.3

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 29.1
1:00 272.7 F 30.0
2:00 273.0 F 29.6
3:00 271.5 F 29.1
4:00 271.5 F 29.4

Min 271.3 F 29.0
Max 273.2 F 30.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 183.4 F 0.0
5:00 220.7 F 0.0
10:00 187.3 F 0.0
15:00 183.3 F 0.0
20:00 178.1 F 0.0
25:00 177.0 F 0.0
30:00 176.6 F 0.0
Drying Complete

FILLING: 0:53
HEATING: 27:41
STERILIZING: 4:00
VENTING: 3:45
DRYING: 30:00
TOTAL CYCLE: 01:06:19
CYCLE COMPLETE

10/25/2024 01:27 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____