

-----  
Midmark M11 - v1.0.0  
Total Cycles: 879

Name: RITTER 1  
Sterilizer ID: V2178198  
06/09/2021 02:33 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 116.5 F 0.0  
2:00 118.6 F 0.0  
4:00 128.3 F 0.4  
6:00 156.8 F 2.4  
8:00 180.8 F 4.7  
10:00 198.7 F 7.3  
12:00 210.1 F 10.4  
14:00 224.1 F 7.2  
16:00 224.9 F 8.1  
18:00 235.0 F 12.3  
20:00 236.1 F 9.4  
22:00 241.3 F 11.0  
24:00 245.8 F 13.0  
26:00 256.5 F 18.6  
28:00 266.1 F 24.5

STERILIZING  
mm:ss Degrees PSI  
0:00 271.2 F 28.2  
1:00 272.1 F 28.9  
2:00 271.9 F 28.4  
3:00 271.0 F 28.2  
4:00 271.8 F 28.9

Min 271.0 F 28.1  
Max 272.3 F 28.9

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 177.4 F 0.0  
5:00 217.2 F 0.0  
10:00 216.6 F 0.0  
15:00 206.6 F 0.0  
20:00 203.9 F 0.0  
25:00 211.1 F 0.0  
30:00 217.8 F 0.0  
Drying Complete

FILLING: 0:60  
HEATING: 29:53  
STERILIZING: 4:00  
VENTING: 3:50  
DRYING: 30:00  
TOTAL CYCLE: 01:08:43  
CYCLE COMPLETE

06/09/2021 03:41 PM

ACCEPTED:

REJECTED:

Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_