
Midmark M11 - v1.0.5
Total Cycles: 3060

Name: RITTER2
Sterilizer ID: V2313228
10/27/2023 03:30 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 109.1 F 0.0
2:00 116.0 F 0.9
4:00 150.6 F 4.3
6:00 175.8 F 9.0
8:00 200.5 F 6.9
10:00 221.3 F 7.5
12:00 236.1 F 12.9
14:00 237.6 F 10.5
16:00 236.2 F 9.1
18:00 251.9 F 16.6
20:00 264.9 F 24.2

STERILIZING
mm:ss Degrees PSI
0:00 272.0 F 28.7
1:00 273.0 F 29.4
2:00 271.8 F 29.0
3:00 272.4 F 29.0
4:00 272.4 F 29.5

Min 271.6 F 28.5
Max 273.0 F 29.6

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 175.5 F 0.0
5:00 216.7 F 0.0
10:00 216.9 F 0.0
15:00 211.4 F 0.0
20:00 206.0 F 0.0
25:00 203.4 F 0.0
30:00 205.7 F 0.0
Drying Complete

FILLING: 1:11
HEATING: 21:32
STERILIZING: 4:00
VENTING: 3:03
DRYING: 30:00
TOTAL CYCLE: 00:59:46
CYCLE COMPLETE

10/27/2023 04:29 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____