

-----  
Midmark M11 - v1.0.0  
Total Cycles: 865

Name: RITTER 1  
Sterilizer ID: V2178198  
06/07/2021 10:22 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 192.0 F 0.0  
2:00 190.7 F 3.9  
4:00 213.9 F 10.4  
6:00 226.7 F 6.6  
8:00 240.4 F 13.5  
10:00 243.9 F 12.2  
12:00 243.3 F 11.7  
14:00 257.2 F 19.2  
16:00 269.4 F 26.9

STERILIZING  
mm:ss Degrees PSI  
0:00 271.1 F 28.4  
1:00 272.2 F 28.6  
2:00 272.9 F 29.1  
3:00 271.4 F 28.3  
4:00 272.6 F 28.9  
  
Min 271.1 F 28.2  
Max 273.0 F 29.2

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 195.4 F 0.0  
5:00 237.1 F 0.0  
10:00 229.4 F 0.0  
15:00 230.4 F 0.0  
20:00 228.4 F 0.0  
25:00 229.2 F 0.0  
30:00 229.3 F 0.0  
Drying Complete

FILLING: 2:30  
HEATING: 16:56  
STERILIZING: 4:00  
VENTING: 3:11  
DRYING: 30:00  
TOTAL CYCLE: 00:56:36  
CYCLE COMPLETE

06/07/2021 11:22 AM  
ACCEPTED: X  
REJECTED:  
Operator: U01