
Midmark M11 - v1.0.0
Total Cycles: 1564

Name: RITTER 1
Sterilizer ID: V2178198
01/26/2022 09:42 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 112.8 F 0.0
2:00 124.9 F 0.3
4:00 134.6 F 2.7
6:00 160.0 F 6.0
8:00 183.0 F 9.9
10:00 204.5 F 6.2
12:00 213.4 F 5.8
14:00 225.4 F 9.9
16:00 233.7 F 9.5
18:00 237.5 F 10.3
20:00 236.5 F 8.5
22:00 246.2 F 13.2
24:00 257.1 F 19.0
26:00 267.2 F 25.2

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 28.2
1:00 272.5 F 29.0
2:00 272.3 F 28.6
3:00 271.4 F 28.1
4:00 271.8 F 28.7

Min 271.3 F 28.1
Max 272.8 F 29.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 188.8 F 0.0
5:00 229.2 F 0.0
10:00 219.0 F 0.0
15:00 220.1 F 0.0
20:00 224.1 F 0.0
25:00 231.5 F 0.0
30:00 224.1 F 0.0
Drying Complete

FILLING: 0:56
HEATING: 27:39
STERILIZING: 4:00
VENTING: 3:14
DRYING: 30:00
TOTAL CYCLE: 01:05:49
CYCLE COMPLETE

01/26/2022 10:47 AM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____