

-----  
Midmark M11 - v1.0.5  
Total Cycles: 3342

Name: RITTER2  
Sterilizer ID: V2313228  
01/31/2024 03:43 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 147.8 F 0.0  
2:00 151.1 F 2.2  
4:00 172.0 F 5.8  
6:00 200.8 F 10.4  
8:00 216.0 F 5.3  
10:00 230.6 F 10.2  
12:00 231.8 F 8.0  
14:00 243.3 F 12.7  
16:00 244.7 F 12.9  
18:00 257.3 F 19.7  
20:00 268.6 F 27.0

STERILIZING  
mm:ss Degrees PSI  
0:00 271.3 F 28.8  
1:00 272.3 F 30.1  
2:00 271.7 F 29.6  
3:00 272.9 F 29.7  
4:00 272.6 F 30.2

Min 270.6 F 28.7  
Max 273.5 F 30.2

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 201.4 F 0.0  
5:00 236.9 F 0.0  
10:00 195.8 F 0.0  
15:00 193.3 F 0.0  
20:00 190.8 F 0.0  
25:00 189.8 F 0.0  
30:00 190.0 F 0.0  
Drying Complete

FILLING: 1:09  
HEATING: 20:58  
STERILIZING: 4:00  
VENTING: 3:29  
DRYING: 30:00  
TOTAL CYCLE: 00:59:36  
CYCLE COMPLETE

01/31/2024 04:42 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_