
Midmark M11 - v1.0.5
Total Cycles: 1449

Name: RITTER2
Sterilizer ID: V2313228
09/02/2022 01:12 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 131.7 F 0.0
2:00 137.6 F 0.0
4:00 156.7 F 1.3
6:00 187.6 F 4.4
8:00 205.9 F 8.7
10:00 221.0 F 12.0
12:00 228.6 F 8.9
14:00 236.8 F 11.4
16:00 241.0 F 11.9
18:00 239.8 F 10.6
20:00 252.9 F 17.2
22:00 264.5 F 24.0

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.8
1:00 272.6 F 29.5
2:00 271.7 F 28.7
3:00 272.3 F 29.6
4:00 272.3 F 29.2

Min 271.4 F 28.6
Max 272.7 F 29.6

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 195.8 F 0.0
5:00 218.1 F 0.0
10:00 230.1 F 0.0
Dry Cycle Aborted

FILLING: 0:60
HEATING: 24:00
STERILIZING: 4:00
VENTING: 3:02
DRYING: 10:15
TOTAL CYCLE: 00:42:16

09/02/2022 01:54 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____