
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.1uS / 0.0ppm
CYCLE NUMBER 001833
10:18 12/07/2024
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 1:58
PRESSURIZING 2:03
STERILIZING 3:04
136.0 C 220kPa 3:04
136.3 C 224kPa 3:34
136.4 C 224kPa 4:04
136.0 C 221kPa 4:34
136.4 C 225kPa 5:04
136.4 C 225kPa 5:34
136.2 C 224kPa 6:04
136.4 C 225kPa 6:34
136.4 C 225kPa 6:34
VENTING 6:34
AIR DRYING 7:22
CYCLE COMPLETE 15:01

Digital Signature #
2218AC4BA313EA66
