

-----  
Midmark M11 - v1.0.0  
Total Cycles: 1558

Name: RITTER 1  
Sterilizer ID: V2178198  
01/25/2022 09:50 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 136.5 F 0.0  
2:00 136.7 F 1.5  
4:00 151.0 F 4.2  
6:00 184.6 F 7.7  
8:00 210.5 F 12.8  
10:00 219.1 F 5.5  
12:00 228.9 F 9.8  
14:00 237.0 F 9.1  
16:00 238.7 F 11.2  
18:00 236.5 F 8.9  
20:00 249.4 F 15.3  
22:00 260.9 F 21.8  
24:00 271.7 F 28.8

STERILIZING  
mm:ss Degrees PSI  
0:00 271.8 F 28.7  
1:00 272.4 F 29.4  
2:00 271.5 F 28.6  
3:00 272.4 F 29.3  
4:00 271.6 F 28.6

Min 271.3 F 28.6  
Max 272.5 F 29.5

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 213.1 F 0.0  
5:00 207.6 F 0.0  
10:00 225.8 F 0.0  
15:00 229.7 F 0.0  
20:00 231.5 F 0.0  
25:00 229.5 F 0.0  
30:00 230.1 F 0.0  
Drying Complete

FILLING: 1:08  
HEATING: 24:27  
STERILIZING: 4:00  
VENTING: 3:14  
DRYING: 30:00  
TOTAL CYCLE: 01:02:49  
CYCLE COMPLETE

01/25/2022 10:52 AM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_