
Midmark M11 - v1.0.0
Total Cycles: 3469

Name: RITTER 1
Sterilizer ID: V2178198
06/30/2023 07:51 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 72.9 F 0.0
2:00 75.6 F 0.0
4:00 96.9 F 0.0
6:00 137.1 F 1.4
8:00 154.3 F 3.0
10:00 176.3 F 5.4
12:00 188.2 F 8.5
14:00 202.8 F 12.1
16:00 211.7 F 3.9
18:00 222.7 F 7.5
20:00 232.8 F 11.3
22:00 230.3 F 6.9
24:00 243.0 F 12.5
26:00 240.2 F 10.4
28:00 252.7 F 16.5
30:00 263.3 F 22.7
32:00 272.8 F 28.6

STERILIZING
mm:ss Degrees PSI
0:00 271.9 F 28.1
1:00 272.2 F 28.6
2:00 272.3 F 28.4
3:00 271.7 F 28.3
4:00 272.8 F 28.8

Min 271.5 F 28.0
Max 272.9 F 28.9

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 173.9 F 0.0
5:00 201.5 F 0.0
10:00 216.3 F 0.0
15:00 221.9 F 0.0
20:00 219.5 F 0.0
25:00 219.5 F 0.0
30:00 218.3 F 0.0

Drying Complete

FILLING: 0:47
HEATING: 32:09
STERILIZING: 4:00
VENTING: 3:09
DRYING: 30:00
TOTAL CYCLE: 01:10:05
CYCLE COMPLETE

06/30/2023 09:01 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____