

-----  
Midmark M11 - v1.0.5  
Total Cycles: 3263

Name: RITTER2  
Sterilizer ID: V2313228  
12/28/2023 04:10 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 185.8 F 0.0  
2:00 181.0 F 3.0  
4:00 194.3 F 8.1  
6:00 214.6 F 10.9  
8:00 227.0 F 9.4  
10:00 234.3 F 9.0  
12:00 244.6 F 13.3  
14:00 244.0 F 12.7  
16:00 257.5 F 19.8  
18:00 269.1 F 27.3

STERILIZING  
mm:ss Degrees PSI  
0:00 271.2 F 28.9  
1:00 272.7 F 29.5  
2:00 271.2 F 28.8  
3:00 272.8 F 29.7  
4:00 271.8 F 29.0  
  
Min 271.1 F 28.7  
Max 272.8 F 29.8

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 179.5 F 0.0  
5:00 228.8 F 0.0  
10:00 192.9 F 0.0  
15:00 188.3 F 0.0  
20:00 187.1 F 0.0  
25:00 185.4 F 0.0  
30:00 183.6 F 0.0  
Drying Complete

FILLING: 1:43  
HEATING: 19:03  
STERILIZING: 4:00  
VENTING: 3:17  
DRYING: 30:00  
TOTAL CYCLE: 00:58:03  
CYCLE COMPLETE

12/28/2023 05:08 PM

ACCEPTED:

REJECTED:

Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_