

-----  
Midmark M11 - v1.0.5  
Total Cycles: 3403

Name: RITTER2  
Sterilizer ID: V2313228  
02/22/2024 02:20 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 103.6 F 0.0  
2:00 104.8 F 0.0  
4:00 114.7 F 0.7  
6:00 144.3 F 3.1  
8:00 172.6 F 6.3  
10:00 193.1 F 10.2  
12:00 208.8 F 7.2  
14:00 219.9 F 8.0  
16:00 233.7 F 12.6  
18:00 232.9 F 9.5  
20:00 237.4 F 9.9  
22:00 245.1 F 13.7  
24:00 257.3 F 20.0  
26:00 268.0 F 26.8

STERILIZING  
mm:ss Degrees PSI  
0:00 271.1 F 29.3  
1:00 272.2 F 29.9  
2:00 271.1 F 29.2  
3:00 272.3 F 30.0  
4:00 271.6 F 29.4

Min 271.0 F 29.2  
Max 272.5 F 30.1

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 186.5 F 0.0  
5:00 204.7 F 0.0  
10:00 209.1 F 0.0  
15:00 208.8 F 0.0  
20:00 206.8 F 0.0  
25:00 206.1 F 0.0  
30:00 204.2 F 0.0  
Drying Complete

FILLING: 0:58  
HEATING: 27:22  
STERILIZING: 4:00  
VENTING: 3:24  
DRYING: 30:00  
TOTAL CYCLE: 01:05:44  
CYCLE COMPLETE

02/22/2024 03:25 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_