
Midmark M11 - v1.0.0
Total Cycles: 4670

Name: RITTER 1
Sterilizer ID: V2178198
06/21/2024 02:47 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 152.1 F 0.0
2:00 162.7 F 2.2
4:00 162.2 F 5.6
6:00 185.1 F 9.4
8:00 209.4 F 10.2
10:00 214.7 F 6.6
12:00 226.3 F 11.2
14:00 228.4 F 6.9
16:00 241.3 F 12.6
18:00 240.4 F 10.8
20:00 252.9 F 17.0
22:00 263.7 F 23.3
24:00 272.0 F 28.7

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 28.5
1:00 272.6 F 29.2
2:00 271.7 F 28.6
3:00 271.7 F 28.9
4:00 272.4 F 29.1

Min 271.2 F 28.4
Max 272.7 F 29.3

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 203.7 F 0.0
5:00 206.2 F 0.0
10:00 209.3 F 0.0
15:00 212.5 F 0.0
20:00 214.4 F 0.0
25:00 221.9 F 0.0
30:00 226.8 F 0.0
Drying Complete

FILLING: 1:02
HEATING: 24:12
STERILIZING: 4:00
VENTING: 3:47
DRYING: 30:00
TOTAL CYCLE: 01:03:01
CYCLE COMPLETE

06/21/2024 03:49 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____