
Midmark M11 - v1.0.5
Total Cycles: 2803

Name: RITTER2
Sterilizer ID: V2313228
08/28/2023 11:12 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 185.5 F 0.0
2:00 173.4 F 0.0
4:00 190.7 F 2.9
6:00 219.4 F 7.5
8:00 224.0 F 12.6
10:00 231.9 F 8.7
12:00 239.1 F 11.0
14:00 243.5 F 12.6
16:00 242.7 F 11.8
18:00 255.5 F 18.5
20:00 266.9 F 25.4

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 28.6
1:00 272.5 F 29.1
2:00 271.4 F 28.8
3:00 272.2 F 29.1
4:00 271.4 F 28.8

Min 271.3 F 28.4
Max 272.8 F 29.4

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 186.6 F 0.0
5:00 226.6 F 0.0
10:00 225.4 F 0.0
15:00 223.1 F 0.0
20:00 230.6 F 0.0
25:00 232.0 F 0.0
30:00 224.5 F 0.0
Drying Complete

FILLING: 0:52
HEATING: 21:29
STERILIZING: 4:00
VENTING: 3:16
DRYING: 30:00
TOTAL CYCLE: 00:59:38
CYCLE COMPLETE

08/28/2023 12:11 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____