

-----  
Midmark M11 - v1.0.0  
Total Cycles: 3669

Name: RITTER 1  
Sterilizer ID: V2178198  
08/17/2023 10:04 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 195.2 F 0.0  
2:00 194.5 F 2.9  
4:00 195.1 F 7.2  
6:00 214.6 F 12.5  
8:00 224.4 F 6.7  
10:00 236.0 F 11.4  
12:00 234.7 F 8.5  
14:00 241.5 F 10.9  
16:00 246.0 F 13.2  
18:00 257.8 F 19.6  
20:00 269.0 F 26.5

STERILIZING  
mm:ss Degrees PSI  
0:00 271.5 F 28.3  
1:00 271.7 F 28.2  
2:00 272.1 F 28.5  
3:00 272.5 F 28.9  
4:00 271.7 F 28.6  
  
Min 271.5 F 28.2  
Max 272.6 F 29.0

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 206.0 F 0.0  
5:00 207.8 F 0.0  
10:00 207.1 F 0.0  
15:00 205.6 F 0.0  
20:00 202.0 F 0.0  
25:00 198.7 F 0.0  
30:00 193.7 F 0.0  
Drying Complete

FILLING: 2:04  
HEATING: 21:02  
STERILIZING: 4:00  
VENTING: 3:22  
DRYING: 30:00  
TOTAL CYCLE: 01:00:28  
CYCLE COMPLETE

08/17/2023 11:04 AM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_