
Midmark M11 - v1.0.5
Total Cycles: 1217

Name: RITTER2
Sterilizer ID: V2313228
06/24/2022 08:34 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 171.5 F 0.0
2:00 166.3 F 2.5
4:00 169.1 F 5.5
6:00 189.8 F 9.8
8:00 214.0 F 6.5
10:00 223.1 F 8.5
12:00 235.8 F 13.5
14:00 237.2 F 10.5
16:00 235.0 F 8.5
18:00 248.7 F 15.1
20:00 260.3 F 21.5
22:00 270.7 F 28.3

STERILIZING
mm:ss Degrees PSI
0:00 271.9 F 29.3
1:00 272.0 F 29.2
2:00 272.0 F 29.3
3:00 272.3 F 29.6
4:00 272.2 F 29.8

Min 271.8 F 29.2
Max 272.6 F 29.8

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 193.8 F 0.0
5:00 219.1 F 0.0
10:00 222.7 F 0.0
15:00 227.1 F 0.0
20:00 220.9 F 0.0
25:00 221.0 F 0.0
30:00 217.9 F 0.0
Drying Complete

FILLING: 1:01
HEATING: 23:32
STERILIZING: 4:00
VENTING: 3:08
DRYING: 30:00
TOTAL CYCLE: 01:01:42
CYCLE COMPLETE

06/24/2022 09:35 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____