
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.3uS / 0.1ppm
CYCLE NUMBER 001003
8:12 16/03/2022
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:41
PRESSURIZING 2:46
STERILIZING 3:54
136.4 C 233kPa 3:54
136.2 C 232kPa 4:24
136.4 C 233kPa 4:54
136.2 C 231kPa 5:24
136.4 C 233kPa 5:54
136.3 C 231kPa 6:24
136.4 C 233kPa 6:54
136.1 C 230kPa 7:24
136.1 C 230kPa 7:24
VENTING 7:24
AIR DRYING 8:14
CYCLE COMPLETE 9:22

Digital Signature #
F044919C1B8DDD6D
