
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.4uS / 0.2ppm
CYCLE NUMBER 001210
11:49 11/08/2022
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:14
PRESSURIZING 3:20
STERILIZING 4:52
136.5 C 225kPa 4:52
136.4 C 225kPa 5:22
136.4 C 225kPa 5:52
136.1 C 223kPa 6:22
136.3 C 224kPa 6:52
136.3 C 224kPa 7:22
136.4 C 225kPa 7:52
136.4 C 225kPa 8:22
136.4 C 225kPa 8:22
VENTING 8:22
AIR DRYING 9:18
CYCLE COMPLETE 12:36

Digital Signature #
81D88752578063D7
