
Midmark M11 - v1.0.5
Total Cycles: 747

Name: RITTER2
Sterilizer ID: V2313228
02/01/2022 03:43 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 124.4 F 0.0
2:00 130.5 F 1.1
4:00 153.6 F 4.3
6:00 181.2 F 8.6
8:00 201.1 F 13.3
10:00 219.7 F 7.3
12:00 233.0 F 12.5
14:00 235.2 F 9.8
16:00 236.8 F 9.0
18:00 248.6 F 14.9
20:00 260.7 F 21.7
22:00 271.3 F 28.9

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 29.0
1:00 272.4 F 29.4
2:00 271.6 F 29.3
3:00 272.3 F 29.4
4:00 271.4 F 29.1

Min 271.3 F 28.9
Max 272.7 F 29.8

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 181.1 F 0.0
5:00 219.2 F 0.0
10:00 201.2 F 0.0
15:00 198.7 F 0.0
20:00 198.6 F 0.0
25:00 197.7 F 0.0
30:00 198.4 F 0.0
Drying Complete

FILLING: 1:07
HEATING: 22:38
STERILIZING: 4:00
VENTING: 3:08
DRYING: 30:00
TOTAL CYCLE: 01:00:53
CYCLE COMPLETE

02/01/2022 04:43 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____