
Midmark M11 - v1.0.0
Total Cycles: 4100

Name: RITTER 1
Sterilizer ID: V2178198
12/08/2023 04:58 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 103.1 F 0.0
2:00 105.1 F 0.9
4:00 119.3 F 3.3
6:00 150.6 F 6.6
8:00 175.9 F 10.3
10:00 211.8 F 5.3
12:00 212.2 F 6.0
14:00 223.2 F 9.9
16:00 233.0 F 9.4
18:00 236.5 F 10.2
20:00 237.1 F 8.9
22:00 245.8 F 13.3
24:00 257.2 F 19.3
26:00 267.5 F 25.7

STERILIZING
mm:ss Degrees PSI
0:00 271.2 F 28.4
1:00 272.5 F 29.1
2:00 271.8 F 28.5
3:00 271.6 F 28.7
4:00 272.7 F 29.1

Min 271.2 F 28.3
Max 272.7 F 29.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 210.2 F 0.0
5:00 225.3 F 0.0
10:00 227.8 F 0.0
15:00 226.7 F 0.0
20:00 225.9 F 0.0
25:00 226.7 F 0.0
30:00 226.3 F 0.0
Drying Complete

FILLING: 1:03
HEATING: 27:28
STERILIZING: 4:00
VENTING: 3:33
DRYING: 30:00
TOTAL CYCLE: 01:06:04
CYCLE COMPLETE

12/08/2023 06:04 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____