

-----  
Midmark M11 - v1.0.5  
Total Cycles: 869

Name: RITTER2  
Sterilizer ID: V2313228  
03/10/2022 08:52 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 168.7 F 0.0  
2:00 167.1 F 2.0  
4:00 185.8 F 6.1  
6:00 208.2 F 11.3  
8:00 223.7 F 6.3  
10:00 237.4 F 11.7  
12:00 236.0 F 9.3  
14:00 238.6 F 9.7  
16:00 248.3 F 14.6  
18:00 260.2 F 21.3  
20:00 271.0 F 28.4

STERILIZING  
mm:ss Degrees PSI  
0:00 271.1 F 28.7  
1:00 272.6 F 29.3  
2:00 271.1 F 28.6  
3:00 272.5 F 29.5  
4:00 272.1 F 29.0

Min 271.0 F 28.6  
Max 272.8 F 29.5

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 189.5 F 0.0  
5:00 231.7 F 0.0  
10:00 235.5 F 0.0  
15:00 225.4 F 0.0  
20:00 218.6 F 0.0  
25:00 226.0 F 0.0  
30:00 200.6 F 0.0  
Drying Complete

FILLING: 2:18  
HEATING: 20:44  
STERILIZING: 4:00  
VENTING: 3:07  
DRYING: 30:00  
TOTAL CYCLE: 01:00:09  
CYCLE COMPLETE

03/10/2022 09:57 AM  
ACCEPTED: X  
REJECTED:  
Operator: U02