
Midmark M11 - v1.0.0
Total Cycles: 4658

Name: RITTER 1
Sterilizer ID: V2178198
06/19/2024 11:11 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 155.9 F 0.0
2:00 154.4 F 2.1
4:00 162.5 F 5.6
6:00 182.4 F 9.5
8:00 199.9 F 10.8
10:00 214.7 F 7.6
12:00 230.9 F 12.2
14:00 232.5 F 8.5
16:00 239.0 F 11.9
18:00 243.0 F 12.0
20:00 253.7 F 18.0
22:00 264.4 F 24.3

STERILIZING
mm:ss Degrees PSI
0:00 272.4 F 29.4
1:00 273.0 F 29.8
2:00 272.3 F 29.8
3:00 272.3 F 29.4
4:00 273.0 F 29.9

Min 271.9 F 29.3
Max 273.2 F 30.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 205.2 F 0.0
5:00 231.9 F 0.0
10:00 228.9 F 0.0
15:00 230.1 F 0.0
20:00 225.2 F 0.0
25:00 228.6 F 0.0
30:00 230.9 F 0.0
Drying Complete

FILLING: 0:59
HEATING: 23:55
STERILIZING: 4:00
VENTING: 3:47
DRYING: 30:00
TOTAL CYCLE: 01:02:41
CYCLE COMPLETE

06/19/2024 12:13 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____