
Midmark M11 - v1.0.0
Total Cycles: 2668

Name: RITTER 1
Sterilizer ID: V2178198
12/13/2022 01:41 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 176.9 F 0.0
2:00 174.0 F 3.1
4:00 186.2 F 7.6
6:00 207.3 F 13.0
8:00 220.8 F 7.9
10:00 236.7 F 13.3
12:00 236.8 F 10.6
14:00 236.8 F 8.8
16:00 250.8 F 16.1
18:00 263.3 F 23.2
20:00 271.7 F 28.5

STERILIZING
mm:ss Degrees PSI
0:00 271.7 F 28.7
1:00 272.7 F 29.4
2:00 271.9 F 28.7
3:00 272.4 F 29.3
4:00 272.6 F 29.3

Min 271.6 F 28.6
Max 272.9 F 29.5

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 208.8 F 0.0
5:00 200.9 F 0.0
10:00 234.7 F 0.0
15:00 244.5 F 0.0
20:00 225.0 F 0.0
25:00 221.4 F 0.0
30:00 220.7 F 0.0
Drying Complete

FILLING: 1:33
HEATING: 20:04
STERILIZING: 4:00
VENTING: 3:13
DRYING: 30:00
TOTAL CYCLE: 00:58:50
CYCLE COMPLETE

12/13/2022 02:39 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____