
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.0uS / 0.0ppm
CYCLE NUMBER 007721
8:16 12/08/2024
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:32
PRESSURIZING 2:38
STERILIZING 3:43
136.0 C 220kPa 3:43
136.4 C 220kPa 4:13
136.2 C 220kPa 4:43
136.3 C 222kPa 5:13
136.4 C 224kPa 5:43
136.2 C 222kPa 6:13
136.4 C 221kPa 6:43
136.3 C 222kPa 7:13
136.3 C 222kPa 7:13
VENTING 7:13
AIR DRYING 8:04
CYCLE COMPLETE 8:23

Digital Signature #
525FBD1A0BBB9031
