
Midmark M11 - v1.0.0
Total Cycles: 3691

Name: RITTER 1
Sterilizer ID: V2178198
08/28/2023 01:32 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 114.6 F 0.0
2:00 117.5 F 1.4
4:00 148.3 F 4.5
6:00 170.8 F 8.0
8:00 194.7 F 12.3
10:00 214.0 F 5.1
12:00 225.8 F 9.2
14:00 234.5 F 11.0
16:00 237.0 F 10.1
18:00 237.3 F 8.7
20:00 247.1 F 13.8
22:00 258.6 F 20.1
24:00 269.6 F 26.8

STERILIZING
mm:ss Degrees PSI
0:00 271.1 F 28.2
1:00 272.2 F 28.9
2:00 271.1 F 28.2
3:00 272.3 F 29.0
4:00 271.5 F 28.3

Min 271.1 F 28.1
Max 272.7 F 29.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 180.3 F 0.0
5:00 230.7 F 0.0
10:00 231.8 F 0.0
15:00 226.7 F 0.0
20:00 207.9 F 0.0
25:00 211.6 F 0.0
30:00 211.1 F 0.0
Drying Complete

FILLING: 0:59
HEATING: 25:02
STERILIZING: 4:00
VENTING: 3:22
DRYING: 30:00
TOTAL CYCLE: 01:03:24
CYCLE COMPLETE

08/28/2023 02:35 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____