
Midmark M11 - v1.0.5
Total Cycles: 2527

Name: RITTER2
Sterilizer ID: V2313228
06/19/2023 07:44 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

C103: HEATUP MODE
STOP PRESSED
CHAMBER TEMP 72F
CHAMBER PRESS 0.0PSI
** ITEMS NOT STERILE **

06/19/2023 07:45 AM