
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.1uS / 0.0ppm
CYCLE NUMBER 002401
10:57 20/12/2024
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:12
PRESSURIZING 2:19
STERILIZING 3:25
136.4 C 221kPa 3:25
136.4 C 226kPa 3:55
136.4 C 224kPa 4:25
136.2 C 223kPa 4:55
136.2 C 224kPa 5:25
136.2 C 223kPa 5:55
136.0 C 223kPa 6:25
136.3 C 224kPa 6:55
136.3 C 224kPa 6:55
VENTING 6:55
AIR DRYING 7:46
CYCLE COMPLETE 18:30

Digital Signature #
90C228831549C4A2
