

-----  
North York, ON  
Canada  
-----

STATIM 2000 S203R750  
SN 500919I00007  
Unit # : 000  
WATER QUALITY  
0.3uS / 0.1ppm  
CYCLE NUMBER 001120  
14:50 13/06/2022  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 4:22  
PRESSURIZING 4:24  
STERILIZING 5:53  
136.0 C 220kPa 5:53  
136.3 C 222kPa 6:23  
136.4 C 224kPa 6:53  
136.1 C 224kPa 7:23  
136.4 C 225kPa 7:53  
136.4 C 225kPa 8:23  
136.3 C 223kPa 8:53  
136.3 C 224kPa 9:23  
136.3 C 224kPa 9:23  
VENTING 9:23  
AIR DRYING 10:26  
CYCLE COMPLETE 25:26  
-----

Digital Signature #  
DF7731B17DD02452  
-----