
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 000287
11:04 09/10/2020
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:54
PRESSURIZING 3:56
STERILIZING 5:03
136.0 C 229kPa 5:03
136.4 C 233kPa 5:33
136.4 C 233kPa 6:03
136.4 C 232kPa 6:33
136.1 C 230kPa 7:03
136.4 C 233kPa 7:33
136.4 C 232kPa 8:03
136.4 C 232kPa 8:33
136.4 C 232kPa 8:33
VENTING 8:33
AIR DRYING 9:25
CYCLE COMPLETE 9:27

Digital Signature #
C14EB474CDB4E78F
