
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 000210
7:29 22/09/2020
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 4:13
PRESSURIZING 4:14
STERILIZING 5:49
136.4 C 232kPa 5:49
136.3 C 232kPa 6:19
136.3 C 232kPa 6:49
136.3 C 232kPa 7:19
136.3 C 232kPa 7:49
136.3 C 232kPa 8:19
136.4 C 233kPa 8:49
136.3 C 233kPa 9:19
136.3 C 233kPa 9:19
VENTING 9:19
AIR DRYING 10:27
CYCLE COMPLETE 13:17

Digital Signature #
25D26F4FA4202F10
