
Midmark M11 - v1.0.0
Total Cycles: 3299

Name: RITTER 1
Sterilizer ID: V2178198
05/29/2023 02:09 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 207.3 F 0.0
2:00 199.4 F 0.1
4:00 212.7 F 5.2
6:00 227.3 F 10.5
8:00 230.9 F 6.3
10:00 239.2 F 12.2
12:00 238.9 F 10.2
14:00 237.7 F 8.6
16:00 251.5 F 15.8
18:00 263.6 F 22.9
20:00 271.5 F 28.0

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 28.1
1:00 272.6 F 28.8
2:00 271.6 F 28.0
3:00 272.5 F 28.9
4:00 272.2 F 28.5

Min 271.4 F 28.0
Max 272.7 F 28.9

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 212.4 F 0.0
5:00 228.3 F 0.0
10:00 239.5 F 0.0
15:00 235.5 F 0.0
20:00 226.4 F 0.0
25:00 220.8 F 0.0
30:00 218.4 F 0.0
Drying Complete

FILLING: 1:32
HEATING: 20:04
STERILIZING: 4:00
VENTING: 3:18
DRYING: 30:00
TOTAL CYCLE: 00:58:54
CYCLE COMPLETE

05/29/2023 03:07 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____