
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R750
SN 500919I00001
Unit # : 000
WATER QUALITY
0.5uS / 0.3ppm
CYCLE NUMBER 000141
15:27 12/03/2020
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 0:32
PRESSURIZING 0:44
STERILIZING 1:53
136.0 C 220kPa 1:53
136.4 C 224kPa 2:23
136.2 C 224kPa 2:53
136.2 C 223kPa 3:23
136.4 C 222kPa 3:53
136.4 C 224kPa 4:23
136.1 C 222kPa 4:53
136.4 C 224kPa 5:23
136.4 C 224kPa 5:23
VENTING 5:23
AIR DRYING 6:15
CYCLE COMPLETE 6:20

Digital Signature #
FF24B8DED325BF03
