
Midmark M11 - v1.0.0
Total Cycles: 2654

Name: RITTER 1
Sterilizer ID: V2178198
12/07/2022 02:44 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 179.0 F 0.0
2:00 179.5 F 3.5
4:00 201.9 F 7.6
6:00 217.8 F 12.8
8:00 225.4 F 7.2
10:00 237.5 F 12.4
12:00 236.6 F 9.8
14:00 238.6 F 8.5
16:00 249.4 F 15.1
18:00 261.8 F 22.2
20:00 272.8 F 29.3

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.5
1:00 272.7 F 29.4
2:00 272.1 F 28.8
3:00 271.6 F 28.7
4:00 272.8 F 29.4

Min 271.5 F 28.4
Max 272.8 F 29.5

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 214.5 F 0.0
5:00 220.9 F 0.0
10:00 234.9 F 0.0
15:00 232.7 F 0.0
20:00 226.7 F 0.0
25:00 224.5 F 0.0
30:00 223.7 F 0.0
Drying Complete

FILLING: 1:11
HEATING: 20:22
STERILIZING: 4:00
VENTING: 3:11
DRYING: 30:00
TOTAL CYCLE: 00:58:43
CYCLE COMPLETE

12/07/2022 03:42 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____