

-----  
Midmark M11 - v1.0.0  
Total Cycles: 4011

Name: RITTER 1  
Sterilizer ID: V2178198  
11/13/2023 11:38 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 204.2 F 0.0  
2:00 198.6 F 3.1  
4:00 202.6 F 7.7  
6:00 215.1 F 13.2  
8:00 226.4 F 8.3  
10:00 240.5 F 12.7  
12:00 239.9 F 11.1  
14:00 237.8 F 9.4  
16:00 251.7 F 16.1  
18:00 263.3 F 22.8  
20:00 271.9 F 28.4

STERILIZING  
mm:ss Degrees PSI  
0:00 271.6 F 28.2  
1:00 271.7 F 28.6  
2:00 272.6 F 28.8  
3:00 271.4 F 28.2  
4:00 271.3 F 28.5  
  
Min 271.0 F 28.1  
Max 272.8 F 29.0

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 209.0 F 0.0  
5:00 231.3 F 0.0  
10:00 231.2 F 0.0  
15:00 229.1 F 0.0  
20:00 229.7 F 0.0  
25:00 229.8 F 0.0  
30:00 227.4 F 0.0  
Drying Complete

FILLING: 1:52  
HEATING: 20:04  
STERILIZING: 4:00  
VENTING: 3:32  
DRYING: 30:00  
TOTAL CYCLE: 00:59:28  
CYCLE COMPLETE

11/13/2023 12:37 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_