
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 001724
13:14 27/03/2024
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:18
PRESSURIZING 2:23
STERILIZING 3:26
136.4 C 224kPa 3:26
136.3 C 224kPa 3:56
136.3 C 224kPa 4:26
136.2 C 224kPa 4:56
136.4 C 226kPa 5:26
136.4 C 226kPa 5:56
136.4 C 222kPa 6:26
136.1 C 223kPa 6:56
136.1 C 223kPa 6:56
VENTING 6:56
AIR DRYING 7:46
CYCLE COMPLETE 7:58

Digital Signature #
BD58651CC515178D
