
Midmark M11 - v1.0.5
Total Cycles: 3746

Name: RITTER2
Sterilizer ID: V2313228
06/28/2024 12:59 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 157.4 F 0.0
2:00 156.2 F 2.2
4:00 164.8 F 5.5
6:00 186.9 F 9.4
8:00 203.8 F 10.7
10:00 216.2 F 7.9
12:00 227.8 F 12.7
14:00 233.8 F 9.6
16:00 237.1 F 9.8
18:00 246.3 F 14.1
20:00 257.6 F 20.3
22:00 268.2 F 26.9

STERILIZING
mm:ss Degrees PSI
0:00 271.0 F 29.4
1:00 273.4 F 30.0
2:00 270.9 F 29.3
3:00 271.7 F 30.1
4:00 272.6 F 30.8

Min 270.9 F 29.3
Max 274.0 F 30.9

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 197.3 F 0.0
5:00 216.3 F 0.0
10:00 199.9 F 0.0
15:00 197.9 F 0.0
20:00 196.2 F 0.0
25:00 193.5 F 0.0
30:00 195.5 F 0.0
Drying Complete

FILLING: 1:07
HEATING: 23:21
STERILIZING: 4:00
VENTING: 3:48
DRYING: 30:00
TOTAL CYCLE: 01:02:17
CYCLE COMPLETE

06/28/2024 02:01 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____