
Midmark M11 - v1.0.5
Total Cycles: 876

Name: RITTER2
Sterilizer ID: V2313228
03/14/2022 09:43 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 175.9 F 0.0
2:00 171.4 F 3.1
4:00 177.5 F 7.4
6:00 200.3 F 12.6
8:00 221.7 F 7.4
10:00 233.5 F 13.0
12:00 239.9 F 10.7
14:00 237.3 F 9.3
16:00 251.6 F 16.2
18:00 263.2 F 23.0
20:00 272.3 F 28.9

STERILIZING
mm:ss Degrees PSI
0:00 271.8 F 29.2
1:00 272.4 F 29.5
2:00 271.8 F 29.4
3:00 272.1 F 29.4
4:00 271.8 F 29.4

Min 271.6 F 29.0
Max 272.8 F 29.8

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 188.7 F 0.0
5:00 231.2 F 0.0
10:00 226.7 F 0.0
15:00 227.0 F 0.0
20:00 227.0 F 0.0
25:00 228.7 F 0.0
30:00 230.1 F 0.0
Drying Complete

FILLING: 1:23
HEATING: 21:01
STERILIZING: 4:00
VENTING: 3:07
DRYING: 30:00
TOTAL CYCLE: 00:59:32
CYCLE COMPLETE

03/14/2022 10:42 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____