
Midmark M11 - v1.0.5
Total Cycles: 330

Name: RITTER2
Sterilizer ID: V2313228
09/02/2021 09:17 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 172.5 F 0.0
2:00 171.8 F 3.0
4:00 193.7 F 8.0
6:00 215.8 F 11.9
8:00 225.2 F 9.6
10:00 234.9 F 9.0
12:00 244.0 F 13.4
14:00 243.6 F 12.6
16:00 256.5 F 19.4
18:00 267.9 F 26.5

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 29.0
1:00 272.3 F 29.4
2:00 271.8 F 29.5
3:00 272.4 F 29.2
4:00 271.9 F 29.7

Min 271.2 F 28.9
Max 272.8 F 29.8

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 184.9 F 0.0
5:00 207.1 F 0.0
10:00 195.2 F 0.0
15:00 193.8 F 0.0
20:00 190.4 F 0.0
25:00 189.3 F 0.0
30:00 189.7 F 0.0
Drying Complete

FILLING: 2:29
HEATING: 19:16
STERILIZING: 4:00
VENTING: 3:07
DRYING: 30:00
TOTAL CYCLE: 00:58:52
CYCLE COMPLETE

09/02/2021 10:15 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____