
Midmark M11 - v1.0.5
Total Cycles: 450

Name: RITTER2
Sterilizer ID: V2313228
10/14/2021 09:44 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 170.2 F 0.0
2:00 172.6 F 3.7
4:00 197.5 F 9.0
6:00 218.3 F 8.4
8:00 231.2 F 10.2
10:00 234.1 F 8.4
12:00 241.3 F 10.9
14:00 247.7 F 14.6
16:00 260.3 F 21.7
18:00 271.8 F 29.3

STERILIZING
mm:ss Degrees PSI
0:00 271.7 F 29.2
1:00 272.0 F 29.1
2:00 272.1 F 29.3
3:00 272.4 F 29.6
4:00 272.2 F 29.8

Min 271.6 F 29.0
Max 272.7 F 29.9

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 198.7 F 0.0
5:00 217.3 F 0.0
10:00 224.0 F 0.0
15:00 215.7 F 0.0
20:00 211.7 F 0.0
25:00 218.0 F 0.0
30:00 218.8 F 0.0
Drying Complete

FILLING: 2:33
HEATING: 18:28
STERILIZING: 4:00
VENTING: 3:15
DRYING: 30:00
TOTAL CYCLE: 00:58:16
CYCLE COMPLETE

10/14/2021 10:46 AM
ACCEPTED: X
REJECTED:
Operator: U02