
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.3uS / 0.1ppm
CYCLE NUMBER 000407
11:06 26/01/2021
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:15
PRESSURIZING 3:16
STERILIZING 4:20
136.0 C 230kPa 4:20
136.4 C 232kPa 4:50
136.4 C 233kPa 5:20
136.4 C 233kPa 5:50
136.4 C 233kPa 6:20
136.4 C 233kPa 6:50
136.4 C 233kPa 7:20
136.4 C 232kPa 7:50
136.4 C 232kPa 7:50
VENTING 7:50
AIR DRYING 8:46
CYCLE COMPLETE 8:47

Digital Signature #
785FC49FB8C53985
