

-----  
North York, ON  
Canada  
-----

STATIM 2000 S203R750  
SN 500919I00007  
Unit # : 000  
WATER QUALITY  
0.4uS / 0.2ppm  
CYCLE NUMBER 000538  
12:06 25/06/2021  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 2:30  
PRESSURIZING 2:35  
STERILIZING 3:47  
136.0 C 230kPa 3:47  
136.1 C 231kPa 4:17  
136.1 C 231kPa 4:47  
136.1 C 230kPa 5:17  
136.2 C 232kPa 5:47  
136.3 C 233kPa 6:17  
136.2 C 232kPa 6:47  
136.3 C 232kPa 7:17  
136.3 C 232kPa 7:17  
VENTING 7:17  
AIR DRYING 8:10  
CYCLE COMPLETE 9:50  
-----

Digital Signature #  
F7D9AF7625ABE2DC  
-----