

-----  
Herzig Eye Institute Ottawa Inc  
phone/email  
150 Bloor St. West, Suite 210  
Toronto ON M5S2X9  
North York, ON M5S2X9  
Canada

6472616554  
david@herzig-eye.com  
-----

STATIM 2000 S203R751  
SN 500919I00001  
Unit # : 000  
WATER QUALITY  
0.2uS / 0.1ppm  
CYCLE NUMBER 004702  
12:31 14/11/2022  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 2:24  
PRESSURIZING 2:30  
STERILIZING 3:13  
136.0 C 220kPa 3:13  
136.4 C 225kPa 3:43  
136.4 C 224kPa 4:13  
136.4 C 222kPa 4:43  
136.4 C 224kPa 5:13  
136.4 C 224kPa 5:43  
136.2 C 222kPa 6:13  
136.4 C 223kPa 6:43  
136.4 C 223kPa 6:43  
VENTING 6:44  
AIR DRYING 7:35  
CYCLE COMPLETE 14:30

-----  
Digital Signature #  
141978EB0F40E882  
-----