

-----  
North York, ON  
Canada  
-----

STATIM 2000 S203R750  
SN 500919I00007  
Unit # : 000  
WATER QUALITY  
0.3uS / 0.1ppm  
CYCLE NUMBER 002168  
12:20 25/10/2024  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 2:02  
PRESSURIZING 2:06  
STERILIZING 3:04  
136.5 C 224kPa 3:04  
136.1 C 222kPa 3:34  
136.4 C 222kPa 4:04  
136.4 C 225kPa 4:34  
136.4 C 224kPa 5:04  
136.4 C 225kPa 5:34  
136.3 C 225kPa 6:04  
136.4 C 226kPa 6:34  
136.4 C 226kPa 6:34  
VENTING 6:34  
AIR DRYING 7:25  
CYCLE COMPLETE 37:25

-----  
Digital Signature #  
5A1EF6078F722853  
-----